

TABLE 3—ESTIMATED ANNUAL THIRD-PARTY DISCLOSURE BURDEN¹

21 CFR Section	Number of respondents	Number of disclosures per respondent	Total annual disclosures	Average burden per disclosure	Total hours
1271.55(a)	1,551	1,422.88	2,206,890	.5 (30 minutes)	1,103,445
1271.60(c) and (d)(2)	1,375	416	572,000	.5 (30 minutes)	286,000
1271.290(c)	1,561	1,324.08	2,066,890	.083 (5 minutes)	171,552
1271.290(f)	1,561	1	1,561	1	1,561
1271.370(b) and (c)	1,561	1,324.08	2,066,890	.25 (15 minutes)	516,723
Total	2,079,281

¹ There are no capital costs or operating and maintenance costs associated with this collection of information.

Dated: August 31, 2016.

Leslie Kux,

Associate Commissioner for Policy.

[FR Doc. 2016–21351 Filed 9–6–16; 8:45 am]

BILLING CODE 4164–01–P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

[Docket No. FDA–2016–N–1486]

Authorization of Emergency Use of an In Vitro Diagnostic Device for Detection of Zika Virus; Availability

AGENCY: Food and Drug Administration, HHS.

ACTION: Notice.

SUMMARY: The Food and Drug Administration (FDA) is announcing the issuance of an Emergency Use Authorization (EUA) (the Authorization) for an in vitro diagnostic device for detection of the Zika virus in response to the Zika virus outbreak in the Americas. FDA issued this Authorization under the Federal Food, Drug, and Cosmetic Act (the FD&C Act) as requested by Viracor-IBT Laboratories, Inc. The Authorization contains, among other things, conditions on the emergency use of the authorized in vitro diagnostic device. The Authorization follows the February 26, 2016, determination by the Secretary of Health and Human Services (HHS) that there is a significant potential for a public health emergency that has a significant potential to affect national security or the health and security of U.S. citizens living abroad and that involves Zika virus. On the basis of such determination, the Secretary of HHS declared on February 26, 2016, that circumstances exist justifying the authorization of emergency use of in vitro diagnostic tests for detection of Zika virus and/or diagnosis of Zika virus infection, subject to the terms of any authorization issued under the

FD&C Act. The Authorization, which includes an explanation of the reasons for issuance, is reprinted in this document.

DATES: The Authorization is effective as of July 19, 2016.

ADDRESSES: Submit written requests for single copies of the EUA to the Office of Counterterrorism and Emerging Threats, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 1, Rm. 4338, Silver Spring, MD 20993–0002. Send one self-addressed adhesive label to assist that office in processing your request or include a fax number to which the Authorization may be sent. See the **SUPPLEMENTARY INFORMATION** section for electronic access to the Authorization.

FOR FURTHER INFORMATION CONTACT:

Carmen Maher, Office of Counterterrorism and Emerging Threats, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 1, Rm. 4347, Silver Spring, MD 20993–0002, 301–796–8510 (this is not a toll free number).

SUPPLEMENTARY INFORMATION:

I. Background

Section 564 of the FD&C Act (21 U.S.C. 360bbb-3) as amended by the Project BioShield Act of 2004 (Pub. L. 108–276) and the Pandemic and All-Hazards Preparedness Reauthorization Act of 2013 (Pub. L. 113–5) allows FDA to strengthen the public health protections against biological, chemical, nuclear, and radiological agents. Among other things, section 564 of the FD&C Act allows FDA to authorize the use of an unapproved medical product or an unapproved use of an approved medical product in certain situations. With this EUA authority, FDA can help assure that medical countermeasures may be used in emergencies to diagnose, treat, or prevent serious or life-threatening diseases or conditions caused by biological, chemical, nuclear, or radiological agents when there are no

adequate, approved, and available alternatives.

Section 564(b)(1) of the FD&C Act provides that, before an EUA may be issued, the Secretary of HHS must declare that circumstances exist justifying the authorization based on one of the following grounds: (1) A determination by the Secretary of Homeland Security that there is a domestic emergency, or a significant potential for a domestic emergency, involving a heightened risk of attack with a biological, chemical, radiological, or nuclear agent or agents; (2) a determination by the Secretary of Defense that there is a military emergency, or a significant potential for a military emergency, involving a heightened risk to U.S. military forces of attack with a biological, chemical, radiological, or nuclear agent or agents; (3) a determination by the Secretary of HHS that there is a public health emergency, or a significant potential for a public health emergency, that affects, or has a significant potential to affect, national security or the health and security of U.S. citizens living abroad, and that involves a biological, chemical, radiological, or nuclear agent or agents, or a disease or condition that may be attributable to such agent or agents; or (4) the identification of a material threat by the Secretary of Homeland Security under section 319F–2 of the Public Health Service (PHS) Act (42 U.S.C. 247d–6b) sufficient to affect national security or the health and security of U.S. citizens living abroad.

Once the Secretary of HHS has declared that circumstances exist justifying an authorization under section 564 of the FD&C Act, FDA may authorize the emergency use of a drug, device, or biological product if the Agency concludes that the statutory criteria are satisfied. Under section 564(h)(1) of the FD&C Act, FDA is required to publish in the **Federal Register** a notice of each authorization, and each termination or revocation of an authorization, and an explanation of the

reasons for the action. Section 564 of the FD&C Act permits FDA to authorize the introduction into interstate commerce of a drug, device, or biological product intended for use when the Secretary of HHS has declared that circumstances exist justifying the authorization of emergency use. Products appropriate for emergency use may include products and uses that are not approved, cleared, or licensed under sections 505, 510(k), or 515 of the FD&C Act (21 U.S.C. 355, 360(k), and 360e) or section 351 of the PHS Act (42 U.S.C. 262). FDA may issue an EUA only if, after consultation with the HHS Assistant Secretary for Preparedness and Response, the Director of the National Institutes of Health, and the Director of the Centers for Disease Control and Prevention (to the extent feasible and appropriate given the applicable circumstances), FDA¹ concludes: (1) That an agent referred to in a declaration of emergency or threat can cause a serious or life-threatening disease or condition; (2) that, based on the totality of scientific evidence available to FDA, including data from adequate and well-controlled clinical trials, if available, it is reasonable to believe that: (A) The product may be effective in diagnosing, treating, or preventing (i) such disease or condition or (ii) a serious or life-threatening disease or condition caused by a product authorized under section

564, approved or cleared under the FD&C Act, or licensed under section 351 of the PHS Act, for diagnosing, treating, or preventing such a disease or condition caused by such an agent, and (B) the known and potential benefits of the product, when used to diagnose, prevent, or treat such disease or condition, outweigh the known and potential risks of the product, taking into consideration the material threat posed by the agent or agents identified in a declaration under section 564(b)(1)(D) of the FD&C Act, if applicable; (3) that there is no adequate, approved, and available alternative to the product for diagnosing, preventing, or treating such disease or condition; and (4) that such other criteria as may be prescribed by regulation are satisfied.

No other criteria for issuance have been prescribed by regulation under section 564(c)(4) of the FD&C Act. Because the statute is self-executing, regulations or guidance are not required for FDA to implement the EUA authority.

II. EUA Request for an In Vitro Diagnostic Device for Detection of the Zika Virus

On February 26, 2016, the Secretary of HHS determined that there is a significant potential for a public health emergency that has a significant potential to affect national security or the health and security of U.S. citizens living abroad and that involves Zika virus. On February 26, 2016, under section 564(b)(1) of the FD&C Act, and on the basis of such determination, the

Secretary of HHS declared that circumstances exist justifying the authorization of emergency use of in vitro diagnostic tests for detection of Zika virus and/or diagnosis of Zika virus infection, subject to the terms of any authorization issued under section 564 of the FD&C Act. Notice of the determination and declaration of the Secretary was published in the **Federal Register** on March 2, 2016 (81 FR 10878). On July 7, 2016, Viracor-IBT Laboratories, Inc. requested, and on July 19, 2016, FDA issued, an EUA for the Zika Virus Real-time RT-PCR test, subject to the terms of the Authorization.

III. Electronic Access

An electronic version of this document and the full text of the Authorization are available on the Internet at <http://www.regulations.gov>.

IV. The Authorization

Having concluded that the criteria for issuance of the Authorization under section 564(c) of the FD&C Act are met, FDA has authorized the emergency use of an in vitro diagnostic device for detection of Zika virus subject to the terms of the Authorization. The Authorization in its entirety (not including the authorized versions of the fact sheets and other written materials) follows and provides an explanation of the reasons for its issuance, as required by section 564(h)(1) of the FD&C Act:

BILLING CODE 4164-01-P

¹ The Secretary of HHS has delegated the authority to issue an EUA under section 564 of the FD&C Act to the Commissioner of Food and Drugs.



DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration
Silver Spring, MD 20993

July 19, 2016

Michael Seymour
Manager of Regulatory Affairs and Quality Assurance
Viracor-IBT Laboratories, Inc.
1001 NW Technology Drive
Lee's Summit, MO 64086

Dear Mr. Seymour:

This letter is in response to your request that the Food and Drug Administration (FDA) issue an Emergency Use Authorization (EUA) for emergency use of Viracor-IBT Laboratories, Inc.'s ("Viracor-IBT") Zika Virus Real-time RT-PCR test for the qualitative detection of RNA from Zika virus in human serum, plasma or urine (collected alongside a patient-matched serum or plasma specimen) from individuals meeting Centers for Disease Control and Prevention (CDC) Zika virus clinical criteria (e.g., clinical signs and symptoms associated with Zika virus infection) and/or CDC Zika virus epidemiological criteria (e.g., history of residence in or travel to a geographic region with active Zika transmission at the time of travel, or other epidemiologic criteria for which Zika virus testing may be indicated). Testing is limited to Viracor-IBT's laboratory in Lee's Summit, MO, or other laboratories designated by Viracor-IBT that are also certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, to perform high complexity tests, pursuant to section 564 of the Federal Food, Drug, and Cosmetic Act (the Act) (21 U.S.C. § 360bbb-3).¹ Assay results are for the identification of Zika viral RNA. Zika viral RNA is generally detectable in these specimens during the acute phase of infection (approximately 7 days in serum, possibly longer in urine, following onset of symptoms, if present).

On February 26, 2016, pursuant to section 564(b)(1)(C) of the Act (21 U.S.C. § 360bbb-3(b)(1)(C)), the Secretary of Health and Human Services (HHS) determined that there is a significant potential for a public health emergency that has a significant potential to affect national security or the health and security of United States citizens living abroad and that involves Zika virus.² Pursuant to section 564(b)(1) of the Act (21 U.S.C. § 360bbb-3(b)(1)), and on the basis of such determination, the Secretary of HHS then declared that circumstances exist justifying the authorization of the emergency use of *in vitro* diagnostic tests for detection

¹ For ease of reference, this letter will refer to "Viracor-IBT's laboratory in Lee's Summit, MO, or other laboratories designated by Viracor-IBT that are also certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, to perform high complexity tests" as "authorized laboratories."

² As amended by the Pandemic and All Hazards Preparedness Reauthorization Act, Pub. L. No. 113-5, under section 564(b)(1)(C) of the Act, the Secretary may make a determination of a public health emergency, or of a significant potential for a public health emergency.

Page 2 – Mr. Seymour, Viracor-IBT Laboratories, Inc.

of Zika virus and/or diagnosis of Zika virus infection, subject to the terms of any authorization issued under 21 U.S.C. § 360bbb-3(a).³

Having concluded that the criteria for issuance of this authorization under section 564(c) of the Act (21 U.S.C. § 360bbb-3(c)) are met, I am authorizing the emergency use of the Zika Virus Real-time RT-PCR test (as described in the Scope of Authorization section of this letter (Section II)) in individuals meeting CDC Zika virus clinical criteria (e.g., clinical signs and symptoms associated with Zika virus infection) and/or CDC Zika virus epidemiological criteria (e.g., history of residence in or travel to a geographic region with active Zika transmission at the time of travel, or other epidemiologic criteria for which Zika virus testing may be indicated) (as described in the Scope of Authorization section of this letter (Section II)) for the detection of Zika virus infection by authorized laboratories, subject to the terms of this authorization.

I. Criteria for Issuance of Authorization

I have concluded that the emergency use of the Zika Virus Real-time RT-PCR test for the detection of Zika virus and diagnosis of Zika virus infection in the specified population meets the criteria for issuance of an authorization under section 564(c) of the Act, because I have concluded that:

1. The Zika virus can cause Zika virus infection, a serious or life-threatening disease or condition to humans infected with the virus;
2. Based on the totality of scientific evidence available to FDA, it is reasonable to believe that the Zika Virus Real-time RT-PCR test, when used with the specified instrument and in accordance with the Scope of Authorization, may be effective in detecting Zika virus and diagnosing Zika virus infection, and that the known and potential benefits of the Zika Virus Real-time RT-PCR test for detecting Zika virus and diagnosing Zika virus infection outweigh the known and potential risks of such product; and
3. There is no adequate, approved, and available alternative to the emergency use of the Zika Virus Real-time RT-PCR test for detecting Zika virus and diagnosing Zika virus infection.⁴

II. Scope of Authorization

I have concluded, pursuant to section 564(d)(1) of the Act, that the scope of this authorization is limited to the use of the authorized Zika Virus Real-time RT-PCR test by authorized laboratories for the detection of RNA from Zika virus and diagnosis of Zika virus infection in individuals meeting CDC Zika virus clinical criteria (e.g., clinical signs and symptoms associated with Zika virus infection) and/or CDC Zika virus epidemiological criteria (e.g., history of residence in or travel to a geographic region with active Zika transmission at the time of travel, or other epidemiologic criteria for which Zika virus testing may be indicated).

³ HHS. *Determination and Declaration Regarding Emergency Use of in Vitro Diagnostic Tests for Detection of Zika Virus and/or Diagnosis of Zika Virus Infection*. 81 Fed. Reg. 10878 (March 2, 2016).

⁴ No other criteria of issuance have been prescribed by regulation under section 564(c)(4) of the Act.

Page 3 – Mr. Seymour, Viracor-IBT Laboratories, Inc.

The Authorized Zika Virus Real-time RT-PCR test

Viracor-IBT's Zika Virus Real-time RT-PCR test is a real-time reverse transcription PCR assay for the qualitative detection of RNA from Zika virus in serum, plasma, urine (collected alongside a patient-matched serum or plasma specimen) and other authorized specimen types.

To perform the Zika Virus Real-time RT-PCR test, samples are first extracted to isolate the Zika virus RNA. Nucleic acids are isolated and purified from the sample using either the bioMerieux NucliSENS easyMag extraction platform with the protocol for total nucleic acid extraction or other authorized extraction methods. An Internal Control sequence is added to the sample prior to extraction and is used as a control for the sample extraction and the amplification reaction.

The purified nucleic acid is first reverse transcribed into cDNAs and amplified using Life Technologies TaqPath™ 1-step RT-qPCR master mix reagent or other authorized ancillary products. In the amplification process, the probe anneals to the specific target sequence located between the forward and reverse primers. The dual-labeled probes include fluorescent dyes and quenchers and specifically detect the presence of Zika virus and Internal Control amplicons during amplification. During the extension phase of the PCR cycle, the 5' nuclease activity of Taq polymerase degrades the probes, causing the reporter dyes to separate from the quencher dyes, generating a fluorescent signal. With each cycle, additional reporter dye molecules are cleaved from their respective probes, increasing the fluorescence intensity. The RT-PCR is performed on the Applied Biosystems® 7500 Real-Time PCR Instrument (Thermo Fisher Scientific) or other authorized instruments.

The Zika Virus Real-time RT-PCR test uses the following materials, or other authorized materials or ancillary products:

- Zika Virus Enzyme Mix
- Zika Virus Primer/Probe Mix
- Zika Virus Internal Control
- Zika Virus Negative Extraction Control
- Zika Virus Positive Extraction Control
- Zika Virus Positive Amplification Curve Controls
- Water (nuclease free)

The Zika Virus Real-time RT-PCR test requires the following control materials, or other authorized control materials, to be included in each run; all controls listed below must generate expected results in order for a test to be considered valid:

- Internal Control
 - The internal control consists of a bacteriophage MS2 that is added to each specimen prior to extraction, is co-purified with each specimen, and is amplified by a specific primers and probe set.
 - The internal control MS2 controls for sample extraction, reverse transcription, amplification and detection and also ensures the absence of non-specific PCR inhibition of a sample.
- No Template Control

Page 4 – Mr. Seymour, Viracor-IBT Laboratories, Inc.

- RNase-, DNase-free water.
 - A no template control is included in each RT-PCR run of specimen extractions to monitor for Zika virus contamination.
- Zika Virus Negative Extraction Control
 - Known negative sample.
 - A negative extraction control is included in each run of specimen extractions to monitor for Zika virus contamination.
- Zika Virus Positive Extraction Control
 - Live Zika whole virus.
 - A positive control is included in each run of specimen extractions to monitor nucleic acid isolation and detection of Zika virus RNA.
- Zika Virus Positive Amplification Curve Controls
 - Modified plasmid with inserted nucleotide regions from the Zika strain KU497555 (Brazil, 2015) – high and low concentrations.
 - The positive amplification curve controls are included in each RT-PCR run of specimen extractions to demonstrate that the anticipated level of sensitivity has been achieved.

To produce a valid run the test controls must meet the performance specifications outlined in the Instructions for Use.

The above described Zika Virus Real-time RT-PCR test, when labeled consistently with the labeling authorized by FDA entitled “Zika Virus Real-time RT-PCR, Viracor-IBT Laboratories, Inc., Instructions for Use” (available at <http://www.fda.gov/MedicalDevices/Safety/EmergencySituations/ucm161496.htm>), which may be revised by Viracor-IBT in consultation with the Division of Microbiology Devices (DMD)/Office of In Vitro Diagnostics and Radiological Health (OIR)/Center for Devices and Radiological Health (CDRH), is authorized to be distributed to and used by authorized laboratories under this EUA, despite the fact that it does not meet certain requirements otherwise required by federal law.

The above described Zika Virus Real-time RT-PCR test is authorized to be accompanied by the following information pertaining to the emergency use, which is authorized to be made available to health care providers, pregnant women, and other patients:

- Fact Sheet for Health Care Providers: Interpreting Viracor-IBT Laboratories, Inc.’s Zika Virus Real-time RT-PCR Test Results
- Fact Sheet for Pregnant Women: Understanding Results from the Viracor-IBT Laboratories, Inc.’s Zika Virus Real-time RT-PCR Test
- Fact Sheet for Patients: Understanding Results from the Viracor-IBT Laboratories, Inc.’s Zika Virus Real-time RT-PCR Test

As described in Section IV below, Viracor-IBT is also authorized to make available additional information relating to the emergency use of the authorized Zika Virus Real-time RT-PCR test that is consistent with, and does not exceed, the terms of this letter of authorization.

Page 5 – Mr. Seymour, Viracor-IBT Laboratories, Inc.

I have concluded, pursuant to section 564(d)(2) of the Act, that it is reasonable to believe that the known and potential benefits of the authorized Zika Virus Real-time RT-PCR test in the specified population, when used for detection of Zika virus and to diagnose Zika virus infection and used consistently with the Scope of Authorization of this letter (Section II), outweigh the known and potential risks of such a product.

I have concluded, pursuant to section 564(d)(3) of the Act, based on the totality of scientific evidence available to FDA, that it is reasonable to believe that the authorized Zika Virus Real-time RT-PCR test may be effective in the detection of Zika virus and diagnosis of Zika virus infection, when used consistently with the Scope of Authorization of this letter (Section II), pursuant to section 564(c)(2)(A) of the Act.

FDA has reviewed the scientific information available to FDA, including the information supporting the conclusions described in Section I above, and concludes that the authorized Zika Virus Real-time RT-PCR test, when used for detection of Zika virus and to diagnose Zika virus infection in the specified population (as described in the Scope of Authorization of this letter (Section II)), meets the criteria set forth in section 564(c) of the Act concerning safety and potential effectiveness.

The emergency use of the authorized Zika Virus Real-time RT-PCR test under this EUA must be consistent with, and may not exceed, the terms of this letter, including the Scope of Authorization (Section II) and the Conditions of Authorization (Section IV). Subject to the terms of this EUA and under the circumstances set forth in the Secretary of HHS's determination described above and the Secretary of HHS's corresponding declaration under section 564(b)(1), the Zika Virus Real-time RT-PCR test described above is authorized to detect Zika virus and diagnose Zika virus infection in individuals meeting CDC Zika virus clinical criteria (e.g., clinical signs and symptoms associated with Zika virus infection) and/or CDC Zika virus epidemiological criteria (e.g., history of residence in or travel to a geographic region with active Zika virus transmission at the time of travel, or other epidemiologic criteria for which Zika virus testing may be indicated).

This EUA will cease to be effective when the HHS declaration that circumstances exist to justify the EUA is terminated under section 564(b)(2) of the Act or when the EUA is revoked under section 564(g) of the Act.

III. Waiver of Certain Requirements

I am waiving the following requirements for the Zika Virus Real-time RT-PCR test during the duration of this EUA:

- Current good manufacturing practice requirements, including the quality system requirements under 21 CFR Part 820 with respect to the design, manufacture, packaging, labeling, storage, and distribution of the Zika Virus Real-time RT-PCR test.
- Labeling requirements for cleared, approved, or investigational devices, including labeling requirements under 21 CFR 809.10 and 21 CFR 809.30, except for the intended use statement (21 CFR 809.10(a)(2), (b)(2)), adequate directions for use (21 U.S.C. 352(f)), (21 CFR 809.10(b)(5), (7), and (8)), any appropriate limitations

Page 6 – Mr. Seymour, Viracor-IBT Laboratories, Inc.

on the use of the device including information required under 21 CFR 809.10(a)(4), and any available information regarding performance of the device, including requirements under 21 CFR 809.10(b)(12).

IV. Conditions of Authorization

Pursuant to section 564 of the Act, I am establishing the following conditions on this authorization:

Viracor-IBT Laboratories, Inc. and Its Authorized Distributor(s)

- A. Viracor-IBT and its authorized distributor(s) will distribute the authorized Zika Virus Real-time RT-PCR test with the authorized labeling, as may be revised by Viracor-IBT Laboratories, Inc. in consultation with DMD/OIR/CDRH, only to authorized laboratories.
- B. Viracor-IBT and its authorized distributor(s) will provide to authorized laboratories the authorized Zika Virus Real-time RT-PCR test Fact Sheet for Health Care Providers, the authorized Zika Virus Real-time RT-PCR test Fact Sheet for Pregnant Women, and the authorized Zika Virus Real-time RT-PCR test Fact Sheet for Patients.
- C. Viracor-IBT and its authorized distributor(s) will make available on their websites the authorized Zika Virus Real-time RT-PCR test Fact Sheet for Health Care Providers, the authorized Zika Virus Real-time RT-PCR test Fact Sheet for Pregnant Women, and the authorized Zika Virus Real-time RT-PCR test Fact Sheet for Patients.
- D. Viracor-IBT and its authorized distributor(s) will inform authorized laboratories and relevant public health authority(ies) of this EUA, including the terms and conditions herein.
- E. Viracor-IBT and its authorized distributor(s) will ensure that the authorized laboratories using the authorized Zika Virus Real-time RT-PCR test have a process in place for reporting test results to health care providers and relevant public health authorities, as appropriate.⁵
- F. Through a process of inventory control, Viracor-IBT and its authorized distributor(s) will maintain records of device usage.
- G. Viracor-IBT and its authorized distributor(s) will collect information on the performance of the test. Viracor-IBT will report to FDA any suspected occurrence of false positive and false negative results and significant deviations from the established performance characteristics of the test of which Viracor-IBT becomes aware.
- H. Viracor-IBT and its authorized distributor(s) are authorized to make available additional information relating to the emergency use of the authorized Zika Virus Real-

⁵ For questions related to reporting Zika test results to relevant public health authorities, it is recommended that Viracor-IBT and authorized laboratories consult with the applicable country, state or territory health department(s). According to CDC, Zika is a nationally notifiable condition. <http://www.cdc.gov/zika/>.

Page 7 – Mr. Seymour, Viracor-IBT Laboratories, Inc.

time RT-PCR test that is consistent with, and does not exceed, the terms of this letter of authorization.

Viracor-IBT Laboratories, Inc.

- I. Viracor-IBT will notify FDA of any authorized distributor(s) of the Zika Virus Real-time RT-PCR test, including the name, address, and phone number of any authorized distributor(s).
- J. Viracor-IBT will provide its authorized distributor(s) with a copy of this EUA, and communicate to its authorized distributor(s) any subsequent amendments that might be made to this EUA and its authorized accompanying materials (e.g., fact sheets, instructions for use).
- K. Viracor-IBT may request changes to the authorized Zika Virus Real-time RT-PCR test Fact Sheet for Health Care Providers, the authorized Zika Virus Real-time RT-PCR test Fact Sheet for Pregnant Women, and the authorized Zika Virus Real-time RT-PCR test Fact Sheet for Patients. Such requests will be made by Viracor-IBT in consultation with, and require concurrence of, DMD/OIR/CDRH.
- L. Viracor-IBT may request the addition of other instruments for use with the authorized Zika Virus Real-time RT-PCR test. Such requests will be made by Viracor-IBT in consultation with, and require concurrence of, DMD/OIR/CDRH.
- M. Viracor-IBT may request the addition of other extraction methods for use with the authorized Zika Virus Real-time RT-PCR test. Such requests will be made by Viracor-IBT in consultation with, and require concurrence of, DMD/OIR/CDRH.
- N. Viracor-IBT may request the addition of other specimen types for use with the authorized Zika Virus Real-time RT-PCR test. Such requests will be made by Viracor-IBT in consultation with, and require concurrence of, DMD/OIR/CDRH.
- O. Viracor-IBT may request the addition of other control materials for use with the authorized Zika Virus Real-time RT-PCR test. Such requests will be made by Viracor-IBT in consultation with, and require concurrence of, DMD/OIR/CDRH.
- P. Viracor-IBT may request the addition of other materials and ancillary products for use with the authorized Zika Virus Real-time RT-PCR test. Such requests will be made by Viracor-IBT in consultation with, and require concurrence of, DMD/OIR/CDRH.
- Q. Viracor-IBT will assess traceability⁶ of the Zika Virus Real-time RT-PCR test with FDA recommended reference material(s). After submission to FDA and DMD/OIR/CDRH's review of and concurrence with the data, Viracor-IBT will update its labeling to reflect the additional testing.
- R. Viracor-IBT will track adverse events and report to FDA under 21 CFR Part 803.

⁶ Traceability refers to tracing analytical sensitivity/reactivity back to a FDA recommended reference material.

Page 8 – Mr. Seymour, Viracor-IBT Laboratories, Inc.

Authorized Laboratories

- S. Authorized laboratories will include with reports of the results of the Zika Virus Real-time RT-PCR test the authorized Fact Sheet for Health Care Providers, the authorized Fact Sheet for Pregnant Women, and the authorized Fact Sheet for Patients. Under exigent circumstances, other appropriate methods for disseminating these Fact Sheets may be used, which may include mass media.
- T. Authorized laboratories will perform the Zika Virus Real-time RT-PCR test on the Applied Biosystems® 7500 Real-Time PCR Instrument or other authorized instruments.
- U. Authorized laboratories will perform the Zika Virus Real-time RT-PCR test using the bioMerieux NucliSENS easyMag extraction platform with the protocol for total nucleic acid extraction or with other authorized extraction methods.
- V. Authorized laboratories will perform the Zika Virus Real-time RT-PCR test on serum, plasma, or urine (collected with a patient-matched serum or plasma specimen) or with other authorized specimen types.
- W. Authorized laboratories will have a process in place for reporting test results to health care providers and relevant public health authorities, as appropriate.⁷
- X. Authorized laboratories will collect information on the performance of the test and report to Viracor-IBT, any suspected occurrence of false positive or false negative results of which they become aware.
- Y. All laboratory personnel using the test should be appropriately trained in RT-PCR techniques and use appropriate laboratory and personal protective equipment when handling this kit, and use the test in accordance with the authorized labeling.

Viracor-IBT Laboratories, Inc., Its Authorized Distributor(s) and Authorized Laboratories

- Z. Viracor-IBT, its authorized distributor(s) and authorized laboratories, will ensure that any records associated with this EUA are maintained until notified by FDA. Such records will be made available to FDA for inspection upon request.

Conditions Related to Advertising and Promotion

- AA. All advertising and promotional descriptive printed matter relating to the use of the authorized Zika Virus Real-time RT-PCR test shall be consistent with the Fact Sheets and authorized labeling, as well as the terms set forth in this EUA and the applicable requirements set forth in the Act and FDA regulations.
- BB. All advertising and promotional descriptive printed matter relating to the use of the authorized Zika Virus Real-time RT-PCR test shall clearly and conspicuously state

⁷ For questions related to reporting Zika test results to relevant public health authorities, it is recommended that Viracor-IBT and authorized laboratories consult with the applicable country, state or territory health department(s). According to CDC, Zika is a nationally notifiable condition. <http://www.cdc.gov/zika/>.

Page 9 — Mr. Seymour, Viracor-IBT Laboratories, Inc.

that:

- This test has not been FDA cleared or approved;
- This test has been authorized by FDA under an EUA for use by authorized laboratories;
- This test has been authorized only for the detection of RNA from Zika virus and diagnosis of Zika virus infection, not for any other viruses or pathogens; and
- This test is only authorized for the duration of the declaration that circumstances exist justifying the authorization of the emergency use of *in vitro* diagnostic tests for detection of Zika virus and/or diagnosis of Zika virus infection under section 564(b)(1) of the Act, 21 U.S.C. § 360bbb-3(b)(1), unless the authorization is terminated or revoked sooner.

No advertising or promotional descriptive printed matter relating to the use of the authorized Zika Virus Real-time RT-PCR test may represent or suggest that this test is safe or effective for the diagnosis of Zika virus infection.

The emergency use of the authorized Zika Virus Real-time RT-PCR test as described in this letter of authorization must comply with the conditions and all other terms of this authorization.

V. Duration of Authorization

This EUA will be effective until the declaration that circumstances exist justifying the authorization of the emergency use of *in vitro* diagnostic tests for detection of Zika virus and/or diagnosis of Zika virus infection is terminated under section 564(b)(2) of the Act or the EUA is revoked under section 564(g) of the Act.

Sincerely,



Robert M. Califf, M.D.
Commissioner of Food and Drugs

Enclosures

Dated: August 31, 2016.

Leslie Kux,

Associate Commissioner for Policy.

[FR Doc. 2016-21353 Filed 9-6-16; 8:45 am]

BILLING CODE 4164-01-C

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

[Docket No. FDA-2016-N-2523]

Request for Comment on the Status of Vinpocetine

AGENCY: Food and Drug Administration, HHS.

ACTION: Notice.

SUMMARY: The Food and Drug Administration (FDA or we) is

requesting comments related to the regulatory status of vinpocetine. Specifically, we request comments on our tentative conclusion that vinpocetine is not a dietary ingredient and is excluded from the definition of dietary supplement in the Federal Food, Drug, and Cosmetic Act (FD&C Act). This action is being taken as part of an administrative proceeding to determine the regulatory status of vinpocetine. All comments submitted by the comment deadline (see **DATES**) will be accepted as part of the official record for this proceeding.

DATES: Submit either electronic or written comments on the notice by November 7, 2016.

ADDRESSES: You may submit comments as follows:

Electronic Submissions

Submit electronic comments in the following way:

- **Federal eRulemaking Portal:** <http://www.regulations.gov>. Follow the instructions for submitting comments. Comments submitted electronically, including attachments, to <http://www.regulations.gov> will be posted to the docket unchanged. Because your comment will be made public, you are solely responsible for ensuring that your comment does not include any confidential information that you or a third party may not wish to be posted, such as medical information, your or anyone else's Social Security number, or confidential business information, such as a manufacturing process. Please note that if you include your name, contact information, or other information that