

Dated: February 21, 2020.

Lowell J. Schiller,

Principal Associate Commissioner for Policy.

[FR Doc. 2020-03860 Filed 2-25-20; 8:45 am]

BILLING CODE 4164-01-P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

[Docket No. FDA-2014-D-1804]

Product Labeling for Laparoscopic Power Morcellators; Draft Guidance for Industry and Food and Drug Administration Staff; Availability

AGENCY: Food and Drug Administration, HHS.

ACTION: Notice of availability.

SUMMARY: The Food and Drug Administration (FDA or Agency) is announcing the availability of the draft guidance entitled “Product Labeling for Laparoscopic Power Morcellators.” This draft guidance proposes updated “Contraindications” and “Warnings” in product labeling information to reflect the state of the science and available technology regarding use of laparoscopic power morcellators (LPMs). This draft guidance is not final nor is it in effect at this time.

DATES: Submit either electronic or written comments on the draft guidance by April 27, 2020 to ensure that the Agency considers your comment on this draft guidance before it begins work on the final version of the guidance.

ADDRESSES: You may submit comments on any guidance at any time as follows:

Electronic Submissions

Submit electronic comments in the following way:

- *Federal eRulemaking Portal:* <https://www.regulations.gov>. Follow the instructions for submitting comments. Comments submitted electronically, including attachments, to <https://www.regulations.gov> will be posted to the docket unchanged. Because your comment will be made public, you are solely responsible for ensuring that your comment does not include any confidential information that you or a third party may not wish to be posted, such as medical information, your or anyone else’s Social Security number, or confidential business information, such as a manufacturing process. Please note that if you include your name, contact information, or other information that identifies you in the body of your comments, that information will be posted on <https://www.regulations.gov>.

- If you want to submit a comment with confidential information that you do not wish to be made available to the public, submit the comment as a written/paper submission and in the manner detailed (see “Written/Paper Submissions” and “Instructions”).

Written/Paper Submissions

Submit written/paper submissions as follows:

- *Mail/Hand Delivery/Courier (for written/paper submissions):* Dockets Management Staff (HFA-305), Food and Drug Administration, 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852.

- For written/paper comments submitted to the Dockets Management Staff, FDA will post your comment, as well as any attachments, except for information submitted, marked and identified, as confidential, if submitted as detailed in “Instructions.”

Instructions: All submissions received must include the Docket No. FDA-2014-D-1804 for “Product Labeling for Laparoscopic Power Morcellators.” Received comments will be placed in the docket and, except for those submitted as “Confidential Submissions,” publicly viewable at <https://www.regulations.gov> or at the Dockets Management Staff between 9 a.m. and 4 p.m., Monday through Friday.

- **Confidential Submissions**—To submit a comment with confidential information that you do not wish to be made publicly available, submit your comments only as a written/paper submission. You should submit two copies total. One copy will include the information you claim to be confidential with a heading or cover note that states “THIS DOCUMENT CONTAINS CONFIDENTIAL INFORMATION.” The Agency will review this copy, including the claimed confidential information, in its consideration of comments. The second copy, which will have the claimed confidential information redacted/blacked out, will be available for public viewing and posted on <https://www.regulations.gov>. Submit both copies to the Dockets Management Staff. If you do not wish your name and contact information to be made publicly available, you can provide this information on the cover sheet and not in the body of your comments and you must identify this information as “confidential.” Any information marked as “confidential” will not be disclosed except in accordance with 21 CFR 10.20 and other applicable disclosure law. For more information about FDA’s posting of comments to public dockets, see 80 FR 56469, September 18, 2015, or access the information at: <https://www.regulations.gov>.

www.govinfo.gov/content/pkg/FR-2015-09-18/pdf/2015-23389.pdf.

Docket: For access to the docket to read background documents or the electronic and written/paper comments received, go to <https://www.regulations.gov> and insert the docket number, found in brackets in the heading of this document, into the “Search” box and follow the prompts and/or go to the Dockets Management Staff, 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852.

You may submit comments on any guidance at any time (see 21 CFR 10.115(g)(5)).

An electronic copy of the guidance document is available for download from the internet. See the **SUPPLEMENTARY INFORMATION** section for information on electronic access to the guidance. Submit written requests for a single hard copy of the draft guidance document entitled “Product Labeling for Laparoscopic Power Morcellators” to the Office of Policy, Guidance and Policy Development, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 5431, Silver Spring, MD 20993-0002. Send one self-addressed adhesive label to assist that office in processing your request.

FOR FURTHER INFORMATION CONTACT: Veronica Price, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 2659, Silver Spring, MD 20993-0002, 301-796-6538.

SUPPLEMENTARY INFORMATION:

I. Background

Following issuance of the 2014 guidance document entitled “Immediately in Effect Guidance Document: Product Labeling for Laparoscopic Power Morcellators,” (<https://www.fda.gov/regulatory-information/search-fda-guidance-documents/immediately-effect-guidance-document-product-labeling-laparoscopic-power-morcellators>) FDA continued to consider new scientific information and the input of stakeholders. Additional scientific information is available that stratifies the risks of an undetected uterine cancer in women with presumed fibroids based on age.

FDA also considered scientific information pertaining to the risk of spreading benign uterine tissue beyond the uterus during gynecologic surgeries when LPMs are used. Parasitic myomas and disseminated peritoneal leiomyomatosis, while benign, have been associated with the need for additional surgery due to symptoms

such as abdominal pain and distension. Finally, FDA considered additional available mitigations for the spread of uterine tissue. Since 2014, FDA has provided marketing authorization for LPM containment systems intended to isolate and contain tissue that is considered benign. These products have been shown, through bench testing and simulated use testing, to contain such tissue during morcellation.

For these reasons, FDA is proposing in this draft guidance to update its recommendations, as originally described in the 2014 guidance document, concerning the content and format of certain labeling information for LPMs. Specifically, FDA is recommending that manufacturers incorporate into the labeling for these devices information providing greater specificity regarding the risks of use as it relates to age, information regarding the risk of spreading benign uterine tissue, and information regarding the use of LPM containment systems.

FDA considered comments received on the final guidance document that appeared in the **Federal Register** of November 25, 2014 (79 FR 70193). FDA revised the guidance as appropriate in response to the comments.

II. Significance of Guidance

This draft guidance is being issued consistent with FDA’s good guidance practices regulation (21 CFR 10.115). The draft guidance, when finalized, will represent the current thinking of FDA on “Product Labeling for Laparoscopic Power Morcellators.” It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations.

III. Electronic Access

Persons interested in obtaining a copy of the draft guidance may do so by downloading an electronic copy from the internet. A search capability for all Center for Devices and Radiological

Health guidance documents is available at <https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/default.htm>. This guidance document is also available at <https://www.regulations.gov>. Persons unable to download an electronic copy of “Product Labeling for Laparoscopic Power Morcellators” may send an email request to CDRH-Guidance@fda.hhs.gov to receive an electronic copy of the document. Please use the document number 1400052 to identify the guidance you are requesting.

IV. Paperwork Reduction Act of 1995

This draft guidance refers to previously approved collections of information. These collections of information are subject to review by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (44 U.S.C. 3501–3521). The collections of information in the following FDA regulations have been approved by OMB as listed in the following table:

21 CFR part	Topic	OMB control No.
807, subpart E	Premarket notification	0910–0120
800, 801, and 809	Medical Device Labeling Regulations	0910–0485
803	Medical Devices; Medical Device Reporting; Manufacturer reporting, importer reporting, user facility reporting, distributor reporting.	0910–0437

Dated: February 20, 2020.
Lowell J. Schiller,
Principal Associate Commissioner for Policy.
 [FR Doc. 2020–03827 Filed 2–25–20; 8:45 am]
BILLING CODE 4164–01–P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Health Resources and Services Administration

Agency Information Collection Activities: Proposed Collection: Public Comment Request; Information Collection Request Title: Shortage Designation Management System

AGENCY: Health Resources and Services Administration (HRSA), Department of Health and Human Services (HHS).
ACTION: Notice.

SUMMARY: In compliance with the requirement for opportunity for public comment on proposed data collection projects of the Paperwork Reduction Act of 1995, HRSA announces plans to submit an Information Collection Request (ICR), described below, to the

Office of Management and Budget (OMB). Prior to submitting the ICR to OMB, HRSA seeks comments from the public regarding the burden estimate, below, or any other aspect of the ICR.

DATES: Comments on this ICR should be received no later than April 27, 2020.

ADDRESSES: Submit your comments to paperwork@hrsa.gov or mail the HRSA Information Collection Clearance Officer, Room 14N136B, 5600 Fishers Lane, Rockville, MD 20857.

FOR FURTHER INFORMATION CONTACT: To request more information on the proposed project or to obtain a copy of the data collection plans and draft instruments, email paperwork@hrsa.gov or call Lisa Wright-Solomon, the HRSA Information Collection Clearance Officer at (301) 443–1984.

SUPPLEMENTARY INFORMATION: When submitting comments or requesting information, please include the information request collection title for reference.

Information Collection Request Title: Shortage Designation Management System OMB No. 0906–0029—Revision.

Abstract: HRSA’s Bureau of Health Workforce is committed to improving the health of the Nation’s underserved communities and vulnerable populations by developing, implementing, evaluating, and refining programs that strengthen the nation’s health workforce. HHS relies on two federal shortage designations to identify and dedicate resources to areas and populations in greatest need of providers: Health Professional Shortage Area (HPSA) designations and Medically Underserved Area/Medically Underserved Population (MUA/P) designations. HPSA designations are geographic areas, population groups, and facilities that are experiencing a shortage of health professionals. The authorizing statute for the National Health Service Corps (NHSC) created HPSAs to fulfill the statutory requirement that NHSC personnel be directed to areas of greatest need. To further differentiate areas of greatest need, HRSA calculates a score for each HPSA. There are three categories of HPSAs based on health discipline: Primary care, dental health, and mental health. Scores range from 1 to 25 for