

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

[Docket No. FDA-2015-D-3787]

Electromagnetic Compatibility of Medical Devices; Draft Guidance for Industry and Food and Drug Administration Staff; Availability

AGENCY: Food and Drug Administration, HHS.

ACTION: Notice of availability.

SUMMARY: The Food and Drug Administration (FDA or Agency) is announcing the availability of the draft guidance entitled “Electromagnetic Compatibility (EMC) of Medical Devices.” FDA has developed this guidance document to describe relevant information that should be provided in a premarket submission (*i.e.*, premarket approval application (PMA), humanitarian device exemption (HDE), premarket notification (510(k)) submission, investigational device exemption (IDE), and De Novo request) to support a claim of electromagnetic compatibility for electrically powered medical devices and medical devices with electrical or electronic functions. This draft guidance is not intended to change current policy, but to provide specific technical information to address the recommendations originally described in the guidance entitled “Information to Support a Claim of Electromagnetic Compatibility (EMC) of Electrically-Powered Medical Devices” published July 11, 2016 (2016 EMC guidance). This draft guidance is not final nor is it in effect at this time.

DATES: Submit either electronic or written comments on the draft guidance by January 19, 2021 to ensure that the Agency considers your comment on this draft guidance before it begins work on the final version of the guidance.

ADDRESSES: You may submit comments on any guidance at any time as follows:

Electronic Submissions

Submit electronic comments in the following way:

- *Federal eRulemaking Portal:* <https://www.regulations.gov>. Follow the instructions for submitting comments. Comments submitted electronically, including attachments, to <https://www.regulations.gov> will be posted to the docket unchanged. Because your comment will be made public, you are solely responsible for ensuring that your comment does not include any confidential information that you or a third party may not wish to be posted, such as medical information, your or

anyone else’s Social Security number, or confidential business information, such as a manufacturing process. Please note that if you include your name, contact information, or other information that identifies you in the body of your comments, that information will be posted on <https://www.regulations.gov>.

- If you want to submit a comment with confidential information that you do not wish to be made available to the public, submit the comment as a written/paper submission and in the manner detailed (see “Written/Paper Submissions” and “Instructions”).

Written/Paper Submissions

Submit written/paper submissions as follows:

- *Mail/Hand Delivery/Courier (for written/paper submissions):* Dockets Management Staff (HFA-305), Food and Drug Administration, 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852.

- For written/paper comments submitted to the Dockets Management Staff, FDA will post your comment, as well as any attachments, except for information submitted, marked and identified, as confidential, if submitted as detailed in “Instructions.”

Instructions: All submissions received must include the Docket No. FDA-2015-D-3787 for “Electromagnetic Compatibility (EMC) of Medical Devices.” Received comments will be placed in the docket and, except for those submitted as “Confidential Submissions,” publicly viewable at <https://www.regulations.gov> or at the Dockets Management Staff between 9 a.m. and 4 p.m., Monday through Friday, 240-402-7500.

- **Confidential Submissions—**To submit a comment with confidential information that you do not wish to be made publicly available, submit your comments only as a written/paper submission. You should submit two copies total. One copy will include the information you claim to be confidential with a heading or cover note that states “THIS DOCUMENT CONTAINS CONFIDENTIAL INFORMATION.” The Agency will review this copy, including the claimed confidential information, in its consideration of comments. The second copy, which will have the claimed confidential information redacted/blacked out, will be available for public viewing and posted on <https://www.regulations.gov>. Submit both copies to the Dockets Management Staff. If you do not wish your name and contact information to be made publicly available, you can provide this information on the cover sheet and not in the body of your comments and you must identify this information as

“confidential.” Any information marked as “confidential” will not be disclosed except in accordance with 21 CFR 10.20 and other applicable disclosure law. For more information about FDA’s posting of comments to public dockets, see 80 FR 56469, September 18, 2015, or access the information at: <https://www.govinfo.gov/content/pkg/FR-2015-09-18/pdf/2015-23389.pdf>.

Docket: For access to the docket to read background documents or the electronic and written/paper comments received, go to <https://www.regulations.gov> and insert the docket number, found in brackets in the heading of this document, into the “Search” box and follow the prompts and/or go to the Dockets Management Staff, 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852, 240-402-7500.

You may submit comments on any guidance at any time (see 21 CFR 10.115(g)(5)).

An electronic copy of the guidance document is available for download from the internet. See the

SUPPLEMENTARY INFORMATION section for information on electronic access to the guidance. Submit written requests for a single hard copy of the draft guidance document entitled “Electromagnetic Compatibility (EMC) of Medical Devices” to the Office of Policy, Guidance and Policy Development, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 5431, Silver Spring, MD 20993-0002; or to the Office of Communication, Outreach and Development, Center for Biologics Evaluation and Research, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 71, Rm. 3128, Silver Spring, MD 20993-0002. Send one self-addressed adhesive label to assist that office in processing your request.

FOR FURTHER INFORMATION CONTACT: Seth J. Seidman, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 1108, Silver Spring, MD 20993-0002, 301-796-2477; or Stephen Ripley, Center for Biologics Evaluation and Research, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 71, Rm. 7301, Silver Spring, MD 20993-0002, 240-402-7911.

SUPPLEMENTARY INFORMATION:

I. Background

FDA has developed this guidance document to describe relevant information that should be provided in a premarket submission (*i.e.*, PMA, HDE

application, 510(k) submission, IDE application, and De Novo requests) to demonstrate EMC for electrically powered medical devices and medical devices with electrical or electronic functions. Typically, the review of EMC information in a submission is based on the risk associated with malfunction or degradation of the medical device under consideration, where malfunction or degradation could be caused by inadequate EMC. The review is also based on the use of appropriate consensus standards. This draft guidance is intended to replace “Information to Support a Claim of Electromagnetic Compatibility (EMC) of Electrically-Powered Medical Devices” published July 11, 2016. This draft guidance is not intended to change current policy, but to provide specific technical information to address the recommendations in the 2016 EMC guidance.

This draft guidance is being issued consistent with FDA’s good guidance practices regulation (21 CFR 10.115). The draft guidance, when finalized, will

represent the current thinking of FDA on the electromagnetic compatibility of medical devices. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations.

II. Electronic Access

Persons interested in obtaining a copy of the draft guidance may do so by downloading an electronic copy from the internet. A search capability for all Center for Devices and Radiological Health guidance documents is available at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/guidance-documents-medical-devices-and-radiation-emitting-products> or from the Center for Biologics Evaluation and Research at <https://www.fda.gov/vaccines-blood-biologics/guidance-compliance-regulatory-information-biologics/biologics-guidances>. This guidance document is also available at <https://www.regulations.gov> and at

<https://www.fda.gov/regulatory-information/search-fda-guidance-documents>. Persons unable to download an electronic copy of “Electromagnetic Compatibility (EMC) of Medical Devices” may send an email request to CDRH-Guidance@fda.hhs.gov to receive an electronic copy of the document. Please include the complete title and the document number 16040 to identify the guidance you are requesting.

III. Paperwork Reduction Act of 1995

While this guidance contains no collection of information, it does refer to previously approved FDA collections of information. Therefore, clearance by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (PRA) (44 U.S.C. 3501–3521) is not required for this guidance. The previously approved collections of information are subject to review by OMB under the PRA. The collections of information in the following FDA regulations and guidance have been approved by OMB as listed in the following table:

21 CFR part or guidance	Topic	OMB control No.
807, subpart E	Premarket notification	0910–0120
814, subparts A through E	Premarket approval	0910–0231
814, subpart H	Humanitarian Device Exemption	0910–0332
812	Investigational Device Exemption	0910–0078
“De Novo Classification Process (Evaluation of Automatic Class III Designation)”.	De Novo classification process	0910–0844
800, 801, and 809	Medical Device Labeling Regulations	0910–0485
803	Medical Devices; Medical Device Reporting; Manufacturer reporting, importer reporting, user facility reporting, distributor reporting.	0910–0437
820	Current Good Manufacturing Practice (CGMP); Quality System (QS) Regulation.	0910–0073

Dated: November 12, 2020.

Lauren K. Roth,

Acting Principal Associate Commissioner for Policy.

[FR Doc. 2020–25306 Filed 11–16–20; 8:45 am]

BILLING CODE 4164–01–P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Health Resources And Services Administration

National Vaccine Injury Compensation Program; List of Petitions Received

AGENCY: Health Resources and Services Administration (HRSA), Department of Health and Human Services (HHS).

ACTION: Notice.

SUMMARY: HRSA is publishing this notice of petitions received under the National Vaccine Injury Compensation

Program (the Program), as required by the Public Health Service (PHS) Act, as amended. While the Secretary of HHS is named as the respondent in all proceedings brought by the filing of petitions for compensation under the Program, the United States Court of Federal Claims is charged by statute with responsibility for considering and acting upon the petitions.

FOR FURTHER INFORMATION CONTACT: For information about requirements for filing petitions, and the Program in general, contact Lisa L. Reyes, Clerk of Court, United States Court of Federal Claims, 717 Madison Place NW, Washington, DC 20005, (202) 357–6400. For information on HRSA’s role in the Program, contact the Director, National Vaccine Injury Compensation Program, 5600 Fishers Lane, Room 08N146B, Rockville, Maryland 20857; (301) 443–6593, or visit our website at: [http://](http://www.hrsa.gov/vaccinecompensation/index.html)

www.hrsa.gov/vaccinecompensation/index.html.

SUPPLEMENTARY INFORMATION: The Program provides a system of no-fault compensation for certain individuals who have been injured by specified childhood vaccines. Subtitle 2 of Title XXI of the PHS Act, 42 U.S.C. 300aa–10 *et seq.*, provides that those seeking compensation are to file a petition with the United States Court of Federal Claims and to serve a copy of the petition to the Secretary of HHS, who is named as the respondent in each proceeding. The Secretary has delegated this responsibility under the Program to HRSA. The Court is directed by statute to appoint special masters who take evidence, conduct hearings as appropriate, and make initial decisions as to eligibility for, and amount of, compensation.

A petition may be filed with respect to injuries, disabilities, illnesses,