

DEPARTMENT OF HEALTH AND HUMAN SERVICES**Food and Drug Administration****21 CFR Part 876**

[Docket No. FDA-2021-N-0285]

Medical Devices; Gastroenterology-Urology Devices; Classification of the Esophageal Tissue Characterization System

AGENCY: Food and Drug Administration, Department of Health and Human Services (HHS).

ACTION: Final amendment; final order.

SUMMARY: The Food and Drug Administration (FDA or we) is classifying the esophageal tissue characterization system into class II (special controls). The special controls that apply to the device type are identified in this order and will be part of the codified language for the esophageal tissue characterization system's classification. We are taking this action because we have determined that classifying the device into class II (special controls) will provide a reasonable assurance of safety and effectiveness of the device. We believe this action will also enhance patients' access to beneficial innovative devices.

DATES: This order is effective December 2, 2021. The classification was applicable on December 23, 2019.

FOR FURTHER INFORMATION CONTACT: Pramodh Kariyawasam, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 2536, Silver Spring, MD 20993-0002, 301-348-1911, Pramodh.Kariyawasam@fda.hhs.gov.

SUPPLEMENTARY INFORMATION:**I. Background**

Upon request, FDA has classified the esophageal tissue characterization system as class II (special controls), which we have determined will provide a reasonable assurance of safety and effectiveness. In addition, we believe this action will enhance patients' access to beneficial innovation, by placing the device into a lower device class than the automatic class III assignment.

The automatic assignment of class III occurs by operation of law and without any action by FDA, regardless of the level of risk posed by the new device. Any device that was not in commercial

distribution before May 28, 1976, is automatically classified as, and remains within, class III and requires premarket approval unless and until FDA takes an action to classify or reclassify the device (see 21 U.S.C. 360c(f)(1)). We refer to these devices as "postamendments devices" because they were not in commercial distribution prior to the date of enactment of the Medical Device Amendments of 1976, which amended the Federal Food, Drug, and Cosmetic Act (FD&C Act).

FDA may take a variety of actions in appropriate circumstances to classify or reclassify a device into class I or II. We may issue an order finding a new device to be substantially equivalent under section 513(i) of the FD&C Act (21 U.S.C. 360c(i)) to a predicate device that does not require premarket approval. We determine whether a new device is substantially equivalent to a predicate device by means of the procedures for premarket notification under section 510(k) of the FD&C Act (21 U.S.C. 360(k)) and part 807 (21 CFR part 807).

FDA may also classify a device through "De Novo" classification, a common name for the process authorized under section 513(f)(2) of the FD&C Act. Section 207 of the Food and Drug Administration Modernization Act of 1997 established the first procedure for De Novo classification (Pub. L. 105-115). Section 607 of the Food and Drug Administration Safety and Innovation Act modified the De Novo application process by adding a second procedure (Pub. L. 112-144). A device sponsor may utilize either procedure for De Novo classification.

Under the first procedure, the person submits a 510(k) for a device that has not previously been classified. After receiving an order from FDA classifying the device into class III under section 513(f)(1) of the FD&C Act, the person then requests a classification under section 513(f)(2). Under the second procedure, rather than first submitting a 510(k) and then a request for classification, if the person determines that there is no legally marketed device upon which to base a determination of substantial equivalence, that person requests a classification under section 513(f)(2) of the FD&C Act. Under either procedure for De Novo classification, FDA is required to classify the device by written order within 120 days. The classification will be according to the criteria under section 513(a)(1) of the FD&C Act. Although the device was

automatically placed within class III, the De Novo classification is considered to be the initial classification of the device.

When FDA classifies a device into class I or II via the De Novo process, the device can serve as a predicate for future devices of that type, including for 510(k)s (see 21 U.S.C. 360c(f)(2)(B)(i)). As a result, other device sponsors do not have to submit a De Novo request or premarket approval application to market a substantially equivalent device (see 21 U.S.C. 360c(i), defining "substantial equivalence"). Instead, sponsors can use the less-burdensome 510(k) process, when necessary, to market their device.

II. De Novo Classification

On December 17, 2018, Diversatek Healthcare Inc. submitted a request for De Novo classification of the Mucosal Integrity Conductivity Test System. FDA reviewed the request in order to classify the device under the criteria for classification set forth in section 513(a)(1) of the FD&C Act.

We classify devices into class II if general controls by themselves are insufficient to provide reasonable assurance of safety and effectiveness, but there is sufficient information to establish special controls that, in combination with the general controls, provide reasonable assurance of the safety and effectiveness of the device for its intended use (see 21 U.S.C. 360c(a)(1)(B)). After review of the information submitted in the request, we determined that the device can be classified into class II with the establishment of special controls. FDA has determined that these special controls, in addition to the general controls, will provide reasonable assurance of the safety and effectiveness of the device.

Therefore, on December 23, 2019, FDA issued an order to the requester classifying the device into class II. In this final order,¹ FDA is codifying the classification of the device by adding 21 CFR 876.1450. We have named the generic type of device esophageal tissue characterization system, and it is identified as a device intended for obtaining measurements of electrical properties within esophageal tissue.

FDA has identified the following risks to health associated specifically with this type of device and the measures required to mitigate these risks in table 1.

¹ FDA notes that the "ACTION" caption for this final order is styled as "Final amendment; final order," rather than "Final order." Beginning in December 2019, this editorial change was made to

indicate that the document "amends" the Code of Federal Regulations. The change was made in accordance with the Office of Federal Register's (OFR) interpretations of the Federal Register Act (44

U.S.C. chapter 15), its implementing regulations (1 CFR 5.9 and parts 21 and 22), and the Document Drafting Handbook.

TABLE 1—ESOPHAGEAL TISSUE CHARACTERIZATION SYSTEM RISKS AND MITIGATION MEASURES

Identified risks	Mitigation measures
Device malfunction related to: <ul style="list-style-type: none"> • Breaking • Fractures • Unintentional separation of components • Inaccurate reading • Failure to sense • Endoscope incompatibility 	Nonclinical performance testing; Shelf life testing; Software verification, validation, and hazard analysis; and Labeling.
Adverse tissue reaction	Biocompatibility evaluation.
Electrical shock and electrical interference from other devices	Electrical safety testing, Electromagnetic compatibility (EMC) testing, and Labeling.
Procedural risks (which may include procedures of endoscopy with sedation).	Labeling.
Infection/cross-contamination	Reprocessing validation, Labeling.

FDA has determined that special controls, in combination with the general controls, address these risks to health and provide reasonable assurance of safety and effectiveness. For a device to fall within this classification, and thus avoid automatic classification in class III, it would have to comply with the special controls named in this final order. The necessary special controls appear in the regulation codified by this order. This device is subject to premarket notification requirements under section 510(k) of the FD&C Act.

III. Analysis of Environmental Impact

The Agency has determined under 21 CFR 25.34(b) that this action is of a type that does not individually or cumulatively have a significant effect on the human environment. Therefore, neither an environmental assessment nor an environmental impact statement is required.

IV. Paperwork Reduction Act of 1995

This final order establishes special controls that refer to previously approved collections of information found in other FDA regulations and guidance. These collections of information are subject to review by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (44 U.S.C. 3501–3521). The collections of information in the guidance document “De Novo Classification Process (Evaluation of Automatic Class III Designation)” have been approved under OMB control number 0910–0844; the collections of information in 21 CFR part 814, subparts A through E, regarding premarket approval, have been approved under OMB control number 0910–0231; the collections of information in part 807, subpart E, regarding premarket notification submissions, have been approved under OMB control number 0910–0120; the collections of information in 21 CFR

part 820, regarding quality system regulation, have been approved under OMB control number 0910–0073; and the collections of information in 21 CFR part 801, regarding labeling, have been approved under OMB control number 0910–0485.

List of Subjects in 21 CFR Part 876

Medical devices.

Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR part 876 is amended as follows:

PART 876—GASTROENTEROLOGY-UROLOGY DEVICES

■ 1. The authority citation for part 876 continues to read as follows:

Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 360L, 371.

■ 2. Add § 876.1450 to subpart B to read as follows:

§ 876.1450 Esophageal tissue characterization system.

(a) *Identification.* An esophageal tissue characterization system is a device intended for obtaining measurements of electrical properties within esophageal tissue.

(b) *Classification.* Class II (special controls). The special controls for this device are:

(1) All patient contacting components of the device must be demonstrated to be biocompatible.

(2) Performance testing must demonstrate the device can accurately measure the designated electrical characteristics.

(3) Mechanical safety testing must demonstrate that the device will withstand forces encountered during use.

(4) Software verification, validation, and hazard analysis must be performed.

(5) Electromagnetic compatibility and electrical safety, mechanical safety, and

thermal safety of the device must be performed.

(6) Performance data must validate the reprocessing instructions for any reusable components of the device.

(7) Labeling must include:

- (i) Specific instructions regarding the proper placement and use of the device;
- (ii) Instructions for reprocessing of any reusable components; and
- (iii) An expiration date for single use components.

Dated: November 26, 2021.

Lauren K. Roth,

Associate Commissioner for Policy.

[FR Doc. 2021–26200 Filed 12–1–21; 8:45 am]

BILLING CODE 4164–01–P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

21 CFR Part 882

[Docket No. FDA–2021–N–0261]

Medical Devices; Neurological Devices; Classification of the Trunk and Limb Electrical Stimulator To Treat Headache

AGENCY: Food and Drug Administration, Department of Health and Human Services (HHS).

ACTION: Final amendment; final order.

SUMMARY: The Food and Drug Administration (FDA or we) is classifying the trunk and limb electrical stimulator to treat headache into class II (special controls). The special controls that apply to the device type are identified in this order and will be part of the codified language for the trunk and limb electrical stimulator to treat headache’s classification. We are taking this action because we have determined that classifying the device into class II (special controls) will provide a reasonable assurance of safety and